

MODIFICATION RECOMMENDED

1146B-01

S E R V I C E N O T E

Supersedes:
None

1146B 100 kHz/100A Current Probe

Serial Numbers: **US53330792 thru US53330952**

All units expected to fail BW measurement accuracy at DC- 100 KHz -3dB spec. Those units having + gain @ at 0.22 dB or @ 15% higher than upper spec (0 dB).

Parts Required:

P/N	Description	Qty.
1146B-FG	Probe - 100 kHz, 100A AC/DC current probe	1

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION: MODIFICATION RECOMMENDED		
ACTION X ON SPECIFIED FAILURE CATEGORY: <input type="checkbox"/> AGREEABLE TIME	STANDARDS LABOR: 0.5 Hours	
LOCATION <input type="checkbox"/> CUSTOMER INSTALLABLE CATEGORY: <input type="checkbox"/> ON-SITE (active On-site contract required) X SERVICE CENTER <input type="checkbox"/> CHANNEL PARTNER	SERVICE X RETURN INVENTORY: <input type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT	USED X RETURN PARTS: <input type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT
AVAILABILITY: PRODUCT'S SUPPORT LIFE	NO CHARGE AVAILABLE UNTIL: May 1, 2019	
<input type="checkbox"/> Calibration Required X Calibration NOT Required	PRODUCT LINE: 1A AUTHOR: MR	
ADDITIONAL INFORMATION: only specific serial numbers US53330792 thru US53330952		

Situation:

All units expected to fail measurement accuracy at DC- 100 KHz -3dB spec. Those units having + gain @ at 0.22 dB or @ 15% higher than upper spec (0 dB).

Solution/Action:

Replace unit with product replacement probe 1146B-FG

Revision History:

Service Note Revision	Date	Author	Reason For Change
01	4/29/2015	Mark Rowley	As published