

34465A-04

# Information Only Service Note

Supersedes:  
NONE

## 34465A Truevolt Digital Multimeter, 6½ Digit

Serial Numbers: ALL

**The Problem** – Option 3446DIGU is no longer available.

### Parts Required:

P/N	Description	Qty.
-----	-------------	------

NONE

### ADMINISTRATIVE INFORMATION

---

Calibration Required  
 Calibration NOT Required

PRODUCT LINE: GM  
AUTHOR: WY

---

ADDITIONAL INFORMATION:

---

**Situation:**

Digitizing and advanced triggering option (DIG) is now enabled as standard feature in 34465A and 34470A digital multimeters.

**Solution/Action:**

All installed base and new customers can enjoy this feature for free by upgrading their instrument (34465A/34470A) firmware to version 3.00 or above.

- Upgrading instrument without the DIG option previously installed to firmware version 3.00 will show the option as “Included” in the instrument’s Help > About window.
- Upgrading instrument with the DIG option previously installed to version 3.00 will continue to show DIG option as installed.

There is no change in the functionality. All other license options must be purchased.

**Revision History:**

Date	Service Note Revision	Author	Reason for Change
04 Aug 2020	01	WanYee	As Published