

34924A-01

**S E R V I C E N O T E**

Supersedes:  
NONE

34924A 70-Channel Reed Multiplexer for 34980A

Serial Numbers: MY43000122-MY43000125  
 MY44001046-MY44001314  
 MY45000004-MY45000006  
 SG45000002-SG45000015

**Intermittent contact resistance with Analog Bus (ABus) relays.**

**Parts Required:**

P/N	Description	Qty.
34924-60001	70-Channel Reed Multiplexer new replacement module	1

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:	
<b>MODIFICATION RECOMMENDED</b>	
ACTION CATEGORY: <input checked="" type="checkbox"/> ON SPECIFIED FAILURE <input type="checkbox"/> AGREEABLE TIME	STANDARDS LABOR: 0.5 Hours
LOCATION CATEGORY: <input type="checkbox"/> CUSTOMER INSTALLABLE <input type="checkbox"/> ON-SITE <input checked="" type="checkbox"/> SERVICE CENTER <input type="checkbox"/> CHANNEL PARTNER	SERVICE INVENTORY: <input checked="" type="checkbox"/> RETURN <input type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT
AVAILABILITY: PRODUCT'S SUPPORT LIFE	USED PARTS: <input type="checkbox"/> RETURN <input type="checkbox"/> SCRAP <input checked="" type="checkbox"/> SEE TEXT
<input type="checkbox"/> Calibration Required <input checked="" type="checkbox"/> Calibration NOT Required	NO CHARGE AVAILABLE UNTIL: 7/1/2014
	PRODUCT LINE: BL AUTHOR: TP
ADDITIONAL INFORMATION: Used parts are to be returned to factory for failure analysis.	

© AGILENT TECHNOLOGIES, INC. 2013  
 PRINTED IN U.S.A.

June 7, 2013  
 Rev. 20



**Situation:**

The ABus relays used in these affected 34924A modules may exhibit intermittent high contact resistance. This can affect two wire resistance measurements and can cause excessive voltage drop in sensitive signal routing applications.

**Solution/Action:**

If these symptoms are observed, contact Agilent Service for a replacement.

**Note:**

- New specifications apply to module replacement with serial number MY52490001/SG52490001 and above. Please refer to the 34980A datasheet in the product website (<http://www.agilent.com/find/34980A>) for latest specifications.
- If you use Y1131A verification software version 1.0, July 2005, please update to the latest version at the Y1131A product website (<http://www.home.agilent.com/en/pd-1383501-pn-Y1131A/verification-and-diagnostic-tools?nid=-33253.778753.00&cc=US&lc=eng>) to test to the latest specifications.

**Revision History:**

Revision Number	Date	Author	Reason For Change
1.0	6/7/13	TP	As Published