

INFORMATION ONLY – DOES NOT COMMUNICATE
A MODIFICATION OR SAFETY CONDITION

3499B-06

S E R V I C E N O T E

Supersedes:
NONE

3499A/B/C

Serial Numbers: [0000A00000 / 9999Z99999]
Firmware revision: 4.02.00

3499A/B/C hangs with CLOS:STAT? or OPEN ALL

Parts Required:

P/N	Description	Qty.
NONE		

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION: INFORMATION ONLY
AUTHOR: woh PRODUCT LINE: PLAR
ADDITIONAL INFORMATION:

© AGILENT TECHNOLOGIES, INC. 2003
PRINTED IN U.S.A.



October 16, 2003

Situation:

When sending a sequence of commands containing OPEN ALL and CLOS::STAT the instrument may hang.

This is a bug in revision 4.0 of the 3499 firmware.

Resolution Workaround

1. Use the command SYST:CPON ALL instead of OPEN ALL
2. Use the command CLOS? Instead of CLOS:STAT?

These two program snippets demonstrate the problem, and show the workaround.

Program 1:

```
CLOS (@100)
```

```
CLOS:STAT? - system hang
```

Work around solution:

```
CLOS (@100)
```

```
CLOS? (@100)
```

Program 2:

```
CLOS (@100)
```

```
OPEN ALL - system hang
```

Work around solution:

```
CLOS (@100)
```

```
SYST: CPON ALL
```

Check this Service Note for future updates.