

## S E R V I C E N O T E

SUPERSEDES: NONE

**37702A Digital Data Tester**

Serial Numbers: 0000U00000 / 9999U99999

**Display flicker not a fault when used in graph storage mode****Duplicate Service Notes:**

37701A-09  
 37702A-03  
 37704A-03  
 37717A-07  
 37717B-04  
 37721A-13  
 37722A-03  
 37724A-07  
 37732A-06

**Situation:**

Some instruments appear to flicker when used in the graph storage mode, independent of storage resolution. When the measurement starts the pattern on the display changes to indicate both the start of gating and start of storage. This display may flicker. This is not a fault. **DO NOT CHANGE THE DISPLAY!**

Replacing the display **WILL NOT** fix the anomaly.

DATE: October 1996

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

**INFORMATION ONLY**

AUTHOR: GS	ENTITY: 1400	ADDITIONAL INFORMATION:
---------------	-----------------	-------------------------