

S E R V I C E N O T E

SUPERSEDES: NONE

37704A Field Portable SONET Test Set**Serial Numbers:** 0000U00000 / 9999U99999**Display flicker not a fault when used in graph storage mode****Duplicate Service Notes:**

37701A-09
 37702A-03
 37704A-03
 37717A-07
 37717B-04
 37721A-13
 37722A-03
 37724A-07
 37732A-06

Situation:

Some instruments appear to flicker when used in the graph storage mode, independent of storage resolution. When the measurement starts the pattern on the display changes to indicate both the start of gating and start of storage. This display may flicker. This is not a fault. **DO NOT CHANGE THE DISPLAY!**

Replacing the display **WILL NOT** fix the anomaly.

DATE: October 1996

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

| | | |
|-------------------|---------------------|-------------------------|
| AUTHOR: GS | ENTITY: 1400 | ADDITIONAL INFORMATION: |
|-------------------|---------------------|-------------------------|