

# 4287A-12

## S E R V I C E N O T E

Supersedes:  
NONE

4287A RF LCR Meter, 1MHz to 3GHz

Serial Numbers: ALL

Use user's 16195B, when performing the 4287A performance test

**Parts Required:**

P/N	Description	Qty.
NONE		

### ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:	
<b>INFORMATION ONLY</b>	
<input checked="" type="checkbox"/> Calibration Required	PRODUCT LINE: WN
<input type="checkbox"/> Calibration NOT Required	AUTHOR: jm
ADDITIONAL INFORMATION:	

© AGILENT TECHNOLOGIES, INC. 2014  
PRINTED IN U.S.A.

February 4, 2014

Rev. 21



**Agilent Technologies**

**Situation:**

When the 4287A user who performs system calibration with his/her 16195B requests repair or regular calibration of the 4287A, service center must perform the 4287A performance test with his/her 16195B.

**Solution/Action:**

Ask user to return his/her 16195B when he/she requests repair or regular calibration.

**Revision History:**

Revision Number	Date	Author	Reason For Change
4287A-12	1-Feb-2014	jm	As published