

S E R V I C E N O T E

SUPERSEDES: NONE

4349B 4-Channel High Resistance Meter

Serial Numbers:JP1KD00295 / JP1KD0601

Modification to prevent the unit from hanging up when improper OPEN correction is performed

To Be Performed By: Agilent-Qualified Personnel

Parts Required:

P/N	Description	Quantity
04349-85071	4349B ROM Rev.01.04	1

Situation:

The 4349B specified above may hang up when the improper OPEN correction displayed any error message is repeated continuously (about 8 times).

Solution / Action:

Replace the ROM to revision 01.04.

DATE: January 1999

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:			
MODIFICATION RECOMMENDED			
ACTION CATEGORY:	<input type="checkbox"/> IMMEDIATELY <input checked="" type="checkbox"/> ON SPECIFIED FAILURE <input type="checkbox"/> AGREEABLE TIME	STANDARDS:	LABOR 0.5 Hours
LOCATION CATEGORY:	<input type="checkbox"/> CUSTOMER INSTALLABLE <input type="checkbox"/> ON-SITE <input checked="" type="checkbox"/> SERVICE CENTER	SERVICE INVENTORY:	<input type="checkbox"/> RETURN <input type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT
AVAILABILITY:	PRODUCT'S SUPPORT LIFE	USED PARTS:	<input type="checkbox"/> RETURN <input type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT
AUTHOR: KW	ENTITY: 3355	AGILENT RESPONSIBLE UNTIL: January 2001	
		ADDITIONAL INFORMATION:	