

RE DEPT

FILE COPY

21 Dec. 1965

DO NOT REMOVE

-hp- MODEL 618B/620A

SIGNAL GENERATORS

Serials Below 338-4136

RESISTOR CHANGE

To improve reliability of -hp- Model 618B/620A Signal Generators, Serials Below 338-4136, change R103 to a fixed composition, 47K ohm $\pm 10\%$, 2 watt resistor, -hp- Stock Number 0693-4731.

F.