

INFORMATION ONLY – DOES NOT COMMUNICATE  
A MODIFICATION OR SAFETY CONDITION

**66309B-04**

**S E R V I C E N O T E**

---

Supersedes:  
NONE

**66309B Dynamic Measurement DC Source**

**Serial Numbers: GB39050311 / GB39050347**

**Low Range Current Measurement is out of Spec**

**Parts Required:**

<b>P/N</b>	<b>Description</b>	<b>Qty.</b>
NONE		

**ADMINISTRATIVE INFORMATION**

SERVICE NOTE CLASSIFICATION: <b>INFORMATION ONLY</b>
AUTHOR: RJM    PRODUCT LINE: 33
ADDITIONAL INFORMATION:

© AGILENT TECHNOLOGIES, INC. 2002  
PRINTED IN U.S.A.



October 17, 2002

**Situation:**

The low current range current shunt in the final test set went out of tolerance, therefore the units which passed through the test set were miss calibrated.

**Solution/Action:**

Follow the section in the User's Guide to calibrate the Low-Range Current Measurement portion of the instrument.