

INFORMATION ONLY – DOES NOT COMMUNICATE
A MODIFICATION OR SAFETY CONDITION

66319B-03

S E R V I C E N O T E

Supersedes:
NONE

66319B Mobile Communications DC Source

Serial Numbers: GB40170326 / GB40170385

Low Range Current Measurement is out of Spec

Parts Required:

P/N	Description	Qty.
NONE		

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION: INFORMATION ONLY
AUTHOR: RJM PRODUCT LINE: 33
ADDITIONAL INFORMATION:

© AGILENT TECHNOLOGIES, INC. 2002
PRINTED IN U.S.A.



October 17, 2002

Situation:

The low current range current shunt in the final test set went out of tolerance, therefore the units which passed through the test set were miss calibrated.

Solution/Action:

Follow the section in the User's Guide to calibrate the Low-Range Current Measurement portion of the instrument.