

6644A-06

S E R V I C E N O T E

Supersedes:
None

6644A DC POWER SUPPLY, 0-60 V, 0-3.5 A, 210 W. GPIB.

Serial Numbers: 0000A-0000 to 3620A-00696

Discontinued Optical Encoder P/N 0960-0912

Situation:

The **Optical Encoder** changed FROM P/N 0960-0822 TO **Encoder** P/N 0960-0912. These two parts are NOT interchangeable

Solution/Action:

When the Optical Encoder **P/N 0960-0822 is** replaced in a model having a serial in the above range, the affected model must have the entire front panel and the display replaced. Old and new front panels and Old and new displays are not compatible or interchangeable.

Note: When replacing the front panel assembly the instrument unit will need to be initialized and re-calibrated.

Available Front Panel Parts:

| P/N | Description | Qty. |
|-----------|-----------------|------|
| 5040-1735 | Keypad | 1 |
| 5065-7075 | LCD | 1 |
| 5080-2261 | LCD Cable | 1 |
| 5060-3400 | Tested assembly | 1 |

ADMINISTRATIVE INFORMATION

| | |
|--|------------------|
| SERVICE NOTE CLASSIFICATION: | |
| INFORMATION ONLY | |
| AUTHOR: CP | PRODUCT LINE: SP |
| ADDITIONAL INFORMATION: Manually Controlled Models IE: (65XA) The Tested Front panel is model specific the replacement tested panel IE: P/N 06XXX-61001 have been Initialized calibration is required. For System Models (66XA) the instrument must be initialized and re-calibrated. | |

© AGILENT TECHNOLOGIES, INC. 2011
PRINTED IN U.S.A.

December 8, 2011

Rev. 18



Agilent Technologies