

S E R V I C E N O T E

SUPERSEDES: None

70001A Mainframe**Serial Numbers:** 0000A00000 / 3634A08481**Power Supply Replacement Kit****To Be Performed By:** Agilent-Qualified Personnel**Parts Required:**

P/N	Description
70001-60112	P/S Kit

Situation:

The power switch in older units was replaced by a newer switch. This change causes the ON/OFF operation to be reversed. This necessitates the replacement of the front door assembly to reflect this change.

Solution/Action:

For units with older power supply's in the above serial number range (prior to P/N 70001-60119), order the Power Supply Replacement Kit P/N 70001-60112.

DATE: May 1999

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR:

DPD

ENTITY:

5300

ADDITIONAL INFORMATION:

