

S E R V I C E N O T E

SUPERSEDES: None

83220A DCS/PCS Test Set**Serial Numbers:** 0000A00000 / 9999Z99999**Recommended calibration period has been extended from 12 months to 24 months****To Be Performed By:** Agilent-Qualified Personnel Customer**Situation:**

The recommended calibration period for the 83220 has been researched by the division. After analysis of the component parts of the instrument it has been decided to increase the period from 12 months to 24 months.

Solution / Action:

This brings it in line with the recommended calibration cycle of the 8922. Therefore if an 8922 and 83220 are part of a system then they will be away for calibration at the same time.

DATE: September 1998

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR:

RS

ENTITY:

E600

ADDITIONAL INFORMATION:

None

