

LOCATION ON-SITE SERVICE □ RETURN USED RETURN CATEGORY: See Text □ SCRAP □ SERVICE CENTER INVENT ORY: SCRAP PARTS: SEETEXT SEETEXT ALWAYS AVAILABILITY: AGILENT RESPONSIBLE UNTIL: ALWAYS AUTHOR:GPB ENTIT Y:5320 ADDITIONAL INFORMATION: None

© 1998 AGILENT TECHNOLOGIES PRINTED IN U.S.A.



8560E-18-S

To Be Performed By: Agilent-Qualified Personnel, Customer

Parts Required:

P/N	Description	Quantity
1540-1130	Soft Carrying Case	1

Situation:

Agilent has learned that the stitching which fastens the shoulder strap to the bag on some soft carrying cases may fail the load test (four times the intended mass). Failure of the stitching can result in damage to the instrument and possibly personal injury.

Solution / Action:

The carrying case which came with option 042 instruments in the serial ranges called out should be sent back to Agilent for exchange. New cases will be supplied under warranty. This exchange also applies to 1540- 1130 Carrying Cases purchased separately from Agilent during the 1997 calendar year. Customers may arrange to have their cases exchanged by contacting the Microwave Instruments Division Spectrum Analyzer Support Line at (707)577-6801.