

S E R V I C E N O T E

SUPERSEDES: None

8564 E Series Portable Spectrum Analyzers

Serial Numbers: See below

A1A2 RPG Assembly and A1 Keyboard Part Number Change

Duplicate Service Notes:

- 8560E-23 Serial Numbers: 0000A00000/3728A03197
- 8561E-19 Serial Numbers: 0000A00000/3728A02040
- 8562E-11 Serial Numbers: 0000A00000/3728A00517
- 8563A-25 Serial Numbers: 0000A00000/3729A04268
- 8563E-23 Serial Numbers: 0000A00000/3728A07704
- 8564E-18 Serial Numbers: 0000A00000/3728A00928
- 8565E-18 Serial Numbers: 0000A00000/3728A00768

To Be Performed By: Agilent-Qualified Personnel, Customer

Parts Required: See Text

Situation:

The original A1A2 Rotary Pulse Generator (RPG), 0960-0745, has been discontinued by the manufacturer. A new RPG, 1990-1525 is used in all instruments with serials above those listed. To accomodate the new RPG, the A1 Keyboard Assembly and the RPG knob were also changed. The old parts are kept on Replacement for older units, so retrofits are not necessary.

Continued

DATE: January 1999

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:		
INFORMATION ONLY		
AUTHOR:	ENTITY:	ADDITIONAL INFORMATION:
GPB	5320	

Solution / Action:

Order replacement part numbers for the RPG, keyboard and knob according to the table below.

Description	P/N for Old Serials	P/N for New Serials
A1A2 RPG	0960-0745	1990-1525
A1 Keyboard Ay	08562-60140	08562-60190
RPG Knob	0370-3069	5041-9610