

MODIFICATION RECOMMENDED –  
CORRECTS MANUFACTURING OR DESIGN DEFECTS

**86107A-01**

# S E R V I C E N O T E

Supersedes:  
NONE

## 86107A Optical /Electrical Plug-in

Serial Numbers: 0000A0000/US42180355

### Memory Error in module

To Be Performed By: Agilent-Qualified Personnel or Customer

#### Parts Required:

P/N	Description	Qty.
86101-60015	Digital Interface Assembly	1

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:			
<b>MODIFICATION RECOMMENDED</b>			
ACTION CATEGORY:	<input type="checkbox"/> IMMEDIATELY <input checked="" type="checkbox"/> ON SPECIFIED FAILURE <input type="checkbox"/> AGREEABLE TIME	STANDARDS:	LABOR: 0.0 Hours
LOCATION CATEGORY:	<input type="checkbox"/> CUSTOMER INSTALLABLE <input type="checkbox"/> ON-SITE <input checked="" type="checkbox"/> SERVICE CENTER	SERVICE INVENTORY:	<input type="checkbox"/> RETURN <input checked="" type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT
AVAILABILITY:	PRODUCT'S SUPPORT LIFE	USED PARTS:	<input type="checkbox"/> RETURN <input checked="" type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT
AUTHOR: bmh	PRODUCT LINE: 8F	AGILENT RESPONSIBLE UNTIL:	EOS
ADDITIONAL INFORMATION:			

© AGILENT TECHNOLOGIES, INC. 2004  
PRINTED IN U.S.A.



January 29, 2004

**Situation:**

Occasionally module memory is corrupted on power up, giving error; “Memory error occurred in \_\_\_\_ module: Try reinstalling module”. Additional circuitry added to digital interface assembly to protect module memory.

**Solution/Action:**

Replace Digital Interface Assembly part number **86101-60010** with **86101-60015** and **recalibrate module.**