

S E R V I C E N O T E

SUPERSEDES: None

86120B Multi-Wavelength Meter

Serial Numbers: 0000A00000 / 9999Z99999

86120 Multi-Wavelength Meter cal cycle extended to 2 years.**Duplicate Service Note:**

86120A-01

Situation:

Based on measured performance of field returns, the calibration cycle of the 86120 family will be increased to two years effective 1 Jul 98.

DATE: July 1999

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR:

DS

ENTITY:

5360

ADDITIONAL INFORMATION: