

## S E R V I C E N O T E

SUPERSEDES: None

**86120B Multi-Wavelength Meter****Serial Numbers:** US00000000 / US99999999**To Be Performed By:** Agilent-Qualified Personnel**Parts Required:**

P/N	Description	Quantity
1150-2108	Board DSP 2MB	1

**Situation:**

The existing DSP board is obsolete and has been replaced by a new part with more memory. For repairs requiring the replacement of the DSP board , use Agilent PN 1150-2108 in place of Agilent PN 1150-5006.

DATE: December 2000

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

**INFORMATION ONLY**AUTHOR:  
DSENTITY:  
5360

ADDITIONAL INFORMATION:

