

S E R V I C E N O T E

SUPERSEDES: 8971C-01

8971C Noise Figure Test Set

Serial Numbers: 0000A00101 / 9999A99999

New A2 Assembly

To Be Performed By: Agilent-Qualified Personnel

Situation:

The preferred and recommended replacement for the A2 microprocessor assembly is 08971-60123. The 08971-60123 microprocessor assembly is shipped without the firmware installed.

Solution/Action:

When ordering the 08971-60123 microprocessor assembly for installation in the 8971C, also order one of the following EPROMs as required for your instrument:

8971C Option 001 Order Part Number 08971-80014

-or-

8971C without 001 Order Part Number 08971-80015

Install the new EPROM at A2U22.

DATE: 15 August 1991

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:		
INFORMATION ONLY		
AUTHOR: FS	ENTITY: 0400	ADDITIONAL INFORMATION: