

DSO90254A-01

S E R V I C E N O T E

Supersedes:
None

DSO/DSA90000A Series Oscilloscopes

Serial Numbers: **MY4748001901 thru MY48230000**
SG4748001901 thru SG48230000

The scopes originally shipped with an 80 GB hard drive. The serial numbers listed above when replacing the internal Hard Drive will need an additional power cable adapter installed to power the replacement 250GB internal Hard Drive. New units shipped with the 250GB hard drives have a new power cable and the adapter is not needed.

Parts Required:

P/N	Description	Qty.
54913-61615	SATA HARD DRIVE CABLE	1

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:	
INFORMATION ONLY	
AUTHOR: MR	PRODUCT LINE: PL1A
ADDITIONAL INFORMATION:	

© AGILENT TECHNOLOGIES, INC. 2008
PRINTED IN U.S.A.

June 26, 2008

Rev. 15



Agilent Technologies

Situation:

The internal replacement hard drive has been replaced with a larger drive and it requires a power adapter cable to operate.

Solution/Action:

Install adapter cable 54913-61615 in series with the current internal hard drive power cable.