

S E R V I C E N O T E

SUPERSEDES: None

2048 Mb/s E1 Line Interface

Serial Numbers: 0000A00000 / 9999Z99999 all units.

Performance Test Result Ammendment

To Be Performed By: Agilent-Qualified Personnel

Situation:

The Test result for tst 7.1 Out Port test is too tight and some units may unnecessarily fail the test.

Solution / Action:

Change the test results as follows:

- 7.1 E1 Out Port Test
- Low Level Nominal 1.4v ± 100 mv
- High Level Nominal 2.4v ± 100 mv

DATE: January 1999

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:		
INFORMATION ONLY		
AUTHOR:	ENTITY:	ADDITIONAL INFORMATION:
DH	Y900	Ph: T2105518 ccmail : doug hayne