E4360A-02

## Information Only Service Note

Supersedes: NONE

## E4360A - Modular Solar Array Simulator

Serial Numbers: Below "MY54000200" (blue/white). "MY54000200/MY99999999" (amber)

The Situation - Blue/white Front Panel LCD Display (below) has been obsoleted.



LCD (blue/white display assembly, p/n 2090-1031) has been discontinued by vendor. Replacement assembly (pictured later) is p/n 5067-6532 (amber in color). Difference in color is unfortunate, but cannot be altered.

## ADMINISTRATIVE INFORMATION

[[]] Calibration Required	PRODUCT LINE: SP
[X] Calibration NOT Required Calibration NOT Required	AUTHOR: JO'C

ADDITIONAL INFORMATION:




## Solution/Action:

1) New LCD display (5067-6532) is a mechanical & functional "drop-in" replacement for the original.



2) 4-pole dip switch on PPMC PCA must have pole #3 repositioned to "off" as shown below:



3) Instrument firmware must be version D.04.01 or newer for proper operation of this display.

Revision History:

Date	Service Note Revision	Author	Reason for Change
25 Nov 2015	01	Jo'C	As Published