

E4982A-01

S E R V I C E N O T E

Supersedes:
NONE

E4982A – LCR Meter, 1 MHz to 3 GHz

Serial Numbers: ALL

Use user's 16195B, when performing the E4982A performance test

Parts Required:

P/N	Description	Qty.
NONE		

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:	
INFORMATION ONLY	
<input checked="" type="checkbox"/> Calibration Required	PRODUCT LINE: WN
<input type="checkbox"/> Calibration NOT Required	AUTHOR: jm
ADDITIONAL INFORMATION:	

© AGILENT TECHNOLOGIES, INC. 2014
PRINTED IN U.S.A.

February 4, 2014

Rev. 21



Agilent Technologies

Situation:

When the E4982A user who performs system calibration with his/her 16195B requests repair or regular calibration of the E4982A, service center must perform the E4982A performance test with his/her 16195B.

Solution/Action:

Ask user to return his/her 16195B when he/she requests repair or regular calibration.

Revision History:

Revision Number	Date	Author	Reason For Change
E4982A-01	1-Feb-2014	jm	As published