

E4982A-10

Information Only Service Note

Supersedes: NONE

E4982A - LCR Meter, 1MHz to 300M/ 500M/ 1G/ 3GHz]

Serial Numbers: MY52500001 - MY52599999, SG52500001 - SG52599999

Manufacturing ID Number: N/A

Default Settings for Administrator Account in Windows 10

Parts Required:

P/N Description Qty.

NONE

ADMINISTRATIVE INFORMATION

[[]] Calibration Required PRODUCT LINE: WN [X] Calibration NOT Required AUTHOR: BS

ADDITIONAL INFORMATION: Default Settings for Administrator Account in Windows 10



Situation:

The default settings for Administrator account in E4982A is different for Windows 10.

Solution/Action:

Listed is the default login username and password for Administrator account for E4982A Windows 10.

Windows Version	Username	Password	Description
Windows 10	Administrator	Keysight4u!	Use for Windows 10 built-in administrator
			account. Password is case-sensitive.

For more information, refer to **E4982A Web Help File**.

Revision History:

Date	Service Note Revision	Author	Reason for Change	
25 Mar 2021	E4982A-10	BS	As Published	