# Information Only Service Note

Supersedes: NONE

## E8257D Analog Signal Generator

Serial Numbers: US00000001-US63049999, MY00000001- MY63049999, SG00000001-SG63049999

The Problem – Instrument that has output board, E8251-60434 replaced with E8251-60887 will have the low phase noise specification impacted.

### Parts Required:

P/N Description Qty. E8251-60887 Output Board 1

#### ADMINISTRATIVE INFORMATION

X Calibration Required PRODUCT LINE: 15
[[]] Calibration NOT Required AUTHOR: PY

ADDITIONAL INFORMATION:



#### Situation:

Replacement of assembly E8251-60434 is due to the depletion of a Life Time Buy part on the board. This is to for assurance of supply (AOS).

#### Solution/Action:

The replacement part for E8251-60434 is E8251-60887. Our engineering team has found the closest possible replacement for the mentioned Life Time Buy part. The new part replaced into the instrument will impact:-

- 1) phase noise specification, both the Absolute and residual SSB on units with Option UNX and UNY.
- 2) Harmonics at 250.0001MHz

For more details, you may refer to the E8257D data sheet labelled for serial prefix  $\geq$  MY/SG6305XXXX. The new specification can be obtain from Keysight.com,

https://www.keysight.com/us/en/assets/7018-01211/data-sheets/5989-0698.pdf

Revision History:

| Date         | Service<br>Note<br>Revision | Author       | Reason for Change |  |
|--------------|-----------------------------|--------------|-------------------|--|
| 12 June 2023 | 01                          | Petrina Yong | As Published      |  |