

# E8267D-18A

# S E R V I C E N O T E

Supersedes:  
E8267D-18

## E8267D SIGNAL GENERATOR

Serial Numbers: MY46520005/MY47220097  
SG47224005/SG47224020

**E8267D found to have the same MAC address that can cause the IP conflict when they are connected to the same local network.**

To Be Performed By: Agilent-Qualified Personnel or Customer

**Parts Required:**

P/N	Description	Qty.
NONE		

### ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:			
<b>MODIFICATION RECOMMENDED</b>			
ACTION CATEGORY:	X IMMEDIATELY <input type="checkbox"/> ON SPECIFIED FAILURE <input type="checkbox"/> AGREEABLE TIME	STANDARDS	LABOR: 0.2 Hours
LOCATION CATEGORY:	X CUSTOMER INSTALLABLE <input type="checkbox"/> ON-SITE <input type="checkbox"/> SERVICE CENTER	SERVICE INVENTORY: <input type="checkbox"/> RETURN <input type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT	USED PARTS: <input type="checkbox"/> RETURN <input type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT
AVAILABILITY:	PRODUCT'S SUPPORT LIFE	NO CHARGE AVAILABLE UNTIL: March 2009	
AUTHOR:	Hong Hee, Ng	PRODUCT LINE: PL15	
ADDITIONAL INFORMATION:			

© AGILENT TECHNOLOGIES, INC. 2008  
PRINTED IN U.S.A.



**Situation:**

Media Access Control (MAC address) is a hardware address that uniquely identifies each node of a network. Each of the E8267D instruments should have a unique MAC address. Due to software bugs, some of the E8267D had been shipped with the same MAC address [00:30:D3:09:1F:AF]. When the E8267D instruments that have the similar MAC address are connected to the same local area network, it causes an IP conflict.

**Solution/Action:**

- (1) To verify the MAC address, press 'Utility', 'GPIB/RS-232 LAN', 'LAN setup'.
- (2) If the MAC address is [00:30:D3:09:1F:AF], please proceed with step 3. Else, the instrument is not affected by MAC address issue.
- (3) Go to this URL: <http://ss.tm.agilent.com/MAC/> .
- (4) Save the MAC address tool.exe to the pc.
- (5) Connect one instrument at a time to the LAN/GPIB.
- (6) Get the IP/GPIB address of the instrument.
- (7) Run the MAC address tool.exe.
- (8) Choose the connection type, either IP or GPIB address that use to connect the instrument to the pc.
- (9) Fill in the IP/GPIB address of the instrument to be updated to the column desired.
- (10) The program will check the S/N of the instrument to determine the new MAC address needed.
- (11) If a new MAC address needed, the program will write the new MAC address to the instrument.
- (12) After the program completed to write the new MAC address, the instrument will power cycle the instrument.
- (13) The new MAC address for the instrument will be display at the word panel of the program.
- (14) After the new MAC address installed, the IP address of the instrument will be changed.