# E8362C-06 <u>S E R V I C E N O T E</u>

Supersedes: NONE

## E8362C PNA Series Microwave Network Analyzers

Serial Numbers: MY49020393, MY49020394, MY49020397

Analyzers with the serial numbers listed above must be returned to Agilent for source attenuator(s) replacement.

Parts Required:P/NDescriptionQty.33325-60012source attenuator1 or 2

### ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:			
MODIFICATION RECOMMENDED			
ACTION CATEGORY:	AGREEABLE TIME	STANDARDS LABOR: 3.0 Hours	
LOCATION CATEGORY:	SERVICE CENTER	SERVICE INVENTORY: SCRAP	USED PARTS: SCRAP
AVAILABILITY:	PRODUCT'S SUPPORT LIFE	NO CHARGE AVAILABLE UNTIL: END OF PRODUCT'S SUPPORT LIFE	
AUTHOR: MF		PRODUCT LINE: WN	
ADDITIONAL INFORMATION: Models affected: N5230C, E8362C, E8363C, N5244A, and N5245A. E8362C-06 created 2/13/12.			

© AGILENT TECHNOLOGIES, INC. 2012 PRINTED IN U.S.A.



#### E8362C-06

#### Page 2 of 2

#### Situation:

Analyzers with the serial numbers listed on the previous page contain source attenuator(s) that may malfunction.

#### Solution/Action:

- Return the PNA to Agilent for source attenuator(s) replacement.
- Refer to the PNA Service Guide chapter 7 for repair and post-repair procedures. To view the online Service Guide, see the numbered list below for instructions.

Refer to the PNA Service Guide for instructions on removing/replacing microcircuits, post-repair adjustments/procedures, and verification/performance tests. To view this Service Guide information, use the following steps:

- 1. Go to <u>http://www.agilent.com</u>.
- 2. Enter your PNA model number (Ex: E8362C) in the Search box and click Search.

#### 3. Click Manuals.

- 4. Scroll down and click on the title/hyperlink for the Service Guide.
- 5. When the PDF of the Service Guide is displayed, use the bookmarks in the left column to find the "Removing & Replacing the ... Source Step Attenuators" section in Chapter 7.
- 6. Follow the instructions for removing and replacing the source attenuator(s).
- 7. Successful completion of the post-repair adjustments/procedures and verification/performance tests will set the analyzer to its pre-repair condition. All pre-repair certification still applies.

**NOTE:** The PNA contains extremely sensitive components that can be ruined if mishandled. Follow instructions carefully when making cable connections, especially wire harness connections.