M9601A-01

Modification Recommended Service Note

Supersedes: NONE

M9601A PXIe Precision Source/Measure Unit, 1.25 MSa/s, 10 fA, 210 V, 315 mA

Serial Numbers: MY59440161-MY59440163, MY59440170, MY59440182, MY59440188-MY59440192, MY59440210, MY59440212, MY59440216, MY59440217, MY59440229, MY59440289, MY59440294, MY59440295, MY59440297, MY59440303-MY59440305, MY60230104-MY60230107, MY60230119, MY60230121

The Problem - The M9601A will not meet published specifications for DC measurement accuracy.

Parts Required:

P/N Description Qty. M9601-65001 High Resolution Source/Measure Unit 1

ADMINISTRATIVE INFORMATION

ACTION CATEGORY:	[[]] ON SPECIFIED FAILURE X AGREEABLE TIME	STANDARDS LABOR: 2.0 Hours
LOCATION CATEGORY:	[]] CUSTOMER INSTALLABLE []] ON-SITE (active On-site contract required) X SERVICE CENTER []] CHANNEL PARTNERS	SERVICE: [[]] RETURN USED [[]] RETURN INVENTORY: [[]] SCRAP X SEE TEXT PARTS: X SCRAP [[]] SEE TEXT
AVAILABILITY	': PRODUCT'S SUPPORT LIFE	NO CHARGE AVAILABLE UNTIL: 31 Dec, 2021
	X Calibration Required [[]] Calibration NOT Required	PRODUCT LINE: 1H AUTHOR: HT

ADDITIONAL INFORMATION:

About the SERVICE INVENTORY: All the service inventory is already replaced with good ones. So, no more actions required.



Situation:

Due to some unexpected behavior of the component used in the product, the M9601A PXIe Precision Source/Measure Unit will not meet published specifications for DC measurement accuracy. Gain adjustment may not be performed correctly, reducing both DC current and voltage accuracy by 0.07% across all ranges.

Solution/Action:

Please contact your nearest Keysight Contact Center at this link, www.keysight.com/find/assist to have your unit to be exchanged. (This product is not customer repairable.)

Revision History:

Revision		Reason for Change	
Dec 1, 2020 01	Hiro Tsunemoto	As Published	