

N1911A-11

Information Only Service Note

Supersedes:
NONE

N1911A P-series, Single Channel Power Meter

Serial Numbers: MY61350001 - MY62010001

The Problem

An alternative power line module has been added into the N1911A Approved Vendor List to mitigate parts supply and delivery issues. The alternative power line module (PN N1913-62700) has been used in units with the serial numbers above instead of PN N1912-61036.

Parts Required:

P/N	Description	Qty.
-----	-------------	------

NONE

ADMINISTRATIVE INFORMATION


Calibration Required
 Calibration NOT Required

PRODUCT LINE: WN
AUTHOR: LSL

ADDITIONAL INFORMATION:

Situation:

The original power line module N1912-61036 is used in the N1912A. Due to part supply chain issues, an alternative module N1913-62700 has been added as a secondary source for the N1912A. The N1913-62700 is slightly larger in size but will have no impact to the power meter’s performance or safety requirements. The Secondary part is also used in N1913A and N1914A power meter.



Solution/Action:

Units installed with the secondary source line module (PN N1913-62700) has passed all the safety assessment tests and meets the performance specifications of the N1912A.

No further action is needed.

Revision History:

Date	Service Note Revision	Author	Reason for Change
11 Feb 2022	01	LSL	As Published