

MODIFICATION RECOMMENDED

N5241A-RECALL-01

S E R V I C E N O T E

Supersedes: NONE

N5241A – Performance Network Analyzer (PNA)

Serial Numbers: MY49411155

Calibrations performed on some 4-Port Performance Network Analyzers by Agilent Technologies between April 01, 2010 and April 07, 2011 show a false pass in the test summary for the calibration coefficients test.

Parts Required:

P/N	Description	Qty.
NONE		

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION: MODIFICATION RECOMMENDED		
ACTION CATEGORY: <input type="checkbox"/> ON SPECIFIED FAILURE <input checked="" type="checkbox"/> AGREEABLE TIME	STANDARDS LABOR: 1.0 Hours	
LOCATION CATEGORY: <input type="checkbox"/> CUSTOMER INSTALLABLE <input type="checkbox"/> ON-SITE <input checked="" type="checkbox"/> SERVICE CENTER <input type="checkbox"/> CHANNEL PARTNER	SERVICE INVENTORY: <input checked="" type="checkbox"/> RETURN <input type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT	USED PARTS: <input type="checkbox"/> RETURN <input type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT
AVAILABILITY: PRODUCT'S SUPPORT LIFE	NO CHARGE AVAILABLE UNTIL: 01-April-2012	
AUTHOR: MW	PRODUCT LINE: **	
ADDITIONAL INFORMATION: Gratis Support need to be chosen as the billing type, while creating the customer order.		

© AGILENT TECHNOLOGIES, INC. 2011
PRINTED IN U.S.A.

April 13, 2011

Rev. 18



Situation:

The Quality issue is caused by a Test-Software defect which overrides the ports 1&2 test result with the test result of ports 3&4 in cases where test points of port 1&2 are either failed or indeterminate. Agilent corrected the PNA Calibration Test Software and installed it on all Test Systems by April 4, 2011.

Solution/Action:

1. Customer, please contact nearest Agilent Customer Care Center and send instrument to Agilent for free calibration.
2. Agilent WCSS Service & Support to calibrate affected products at Agilent's expense.
3. After re-calibration Agilent to send the instrument back to the customer.