#### N5245A-05 S V Ν Ε R С Ε Ε ()

Supersedes: None

# N5245A - PNA Series Microwave Network Analyzers

Serial Numbers: MY49151216

Analyzers with the serial numbers listed above were shipped with a front panel interface board that will eventually malfunction.

**Parts Required:** Description P/N Qty. N5240-60065 front panel interface board 1

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:			
MODIFICATION RECOMMENDED			
ACTION CATEGORY:	AGREEABLE TIME	STANDARDS LABOR: 1.0 Hour	
LOCATION CATEGORY:	ON-SITE, or SERVICE CENTER	SERVICE INVENTORY: SCRAP	USED PARTS: SCRAP
AVAILABILITY:	PRODUCT'S SUPPORT LIFE	NO CHARGE AVAILABLE UNTIL: June 30, 2013	
Calibration NOT Required		PRODUCT LINE: WN AUTHOR: MF	
ADDITIONAL INFORMATION: Models affected: N5241A, N5242A, N5242AS, N5244AS, and N5245A. Service Note N5245A-05 created 6/9/2012.			

© AGILENT TECHNOLOGIES, INC. 2012 PRINTED IN U.S.A.



June 11, 2012

### Situation:

Analyzers with the serial numbers listed on the previous page were shipped with a front panel interface board that will eventually malfunction.

### Solution/Action:

Replace the front panel interface board. Refer to the PNA Service Guide chapter 7 section titled, "Removing and Replacing the A1 – A3 and Other Front Panel Subassemblies" for instructions on removing/replacing the assembly, post-repair adjustments/procedures, and verification/performance tests. To view this Service Guide information, use the following steps:

- 1. Go to http://www.agilent.com.
- 2. Enter your PNA model number (Ex: N5245A) in the Search box and click Search.
- 3. Click Manuals.
- 4. Scroll down and click on the title/hyperlink for the Service Guide.
- 5. When the PDF of the Service Guide is displayed, use the bookmarks in the left column to find the "Removing and Replacing the A1 A3 and Other Front Panel Subassemblies" section in Chapter 7.
- 6. Follow the instructions for removing and replacing the assembly.
- 7. Successful completion of the post-repair adjustments/procedures and verification/performance tests will set the analyzer to its pre-repair condition. All pre-repair certification still applies.