N5743A-03 <u>S E R V I C E N O T E</u>

Supersedes: NONE

N5743A DC System Power Supply, 12.5V, 60A, 750W

Serial Numbers: All Units

During verification performance testing of this power supply it can have a unexpected voltage level on its output as result of the verification procedure ending with the maximum voltage setting and not instructing the user to set voltage levels to zero at the end of the verification procedure

Parts Required: P/N Description

Qty.

NONE

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:			
MODIFICATION RECOMMENDED			
ACTION [[]] ON SPECIFIED FAILURE CATEGORY: X AGREEABLE TIME	STANDARDS LABOR: 0.1 Hours		
LOCATION X CUSTOMER INSTALLABLE CATEGORY: [[]] ON-SITE (active On-site contract required) [[]] SERVICE CENTER [[]] CHANNEL PARTNER	SERVICE [[]] RETURN INVENTORY: [[]] SCRAP X SEE TEXT	USED [[]] RETURN PARTS: [[]] SCRAP X SEE TEXT	
AVAILABILITY: ALWAYS	NO CHARGE AVAILABLE UNTIL: EOS		
[[]] Calibration Required X Calibration NOT Required	PRODUCT LINE: SP AUTHOR: PF		
ADDITIONAL INFORMATION: Customer can view latest manual changes at www.agilent.com/find/n5700			

© AGILENT TECHNOLOGIES, INC. 2013 PRINTED IN U.S.A.



July 16, 2013 Rev. 21

Page 2 of 2

Situation:

When performing verification of power supply the last performance test instructs user to set the voltage setting to maximum rating. If power supply is set to auto restart mode it restores the operating setting that was set when it was last turned off. When this occurs the voltage or current will be set to maximum rating the next time it is turned on. This will present an unexpected voltage/current setting condition .

Solution/Action:

- 1. During verification of instrument (s) use the latest version of users guide.
- 2. Download the latest user's guide. Use web link to download latest manual found on Document Library tab. <u>http://www.agilent.com/find/n5700</u>

Revision History:

Revision Number	Date	Author	Reason For Change
1.0	7/11/13	PF	As published