

N9010A-03B

Information Only Service Note

Supersedes:
N9010A-03A

N9010A EXA Signal Analyzer

Serial Numbers: ALL

EXA Signal Analyzer Removable Disk Drive and CPU Compatibility

As of December 2016, there are six versions of CPUs used in EXA signal analyzers that accept removable disk drives, but not all removable disk drives can be used on all versions of CPU. Also, there are now three different Operating Systems.

Parts Required: NONE

ADMINISTRATIVE INFORMATION

Calibration Required
 Calibration NOT Required

PRODUCT LINE: 12
AUTHOR: BDT

ADDITIONAL INFORMATION:

Situation:

As of December, 2016, there are six versions of CPUs used in EXA signal analyzers that accept removable disk drives, but not all removable disk drives can be used on all versions of CPU. It is the image with which the disk drive was built that determines with what CPUs the disk drive is compatible. The currently-running instrument software version is not necessarily the same as the image with which the disk drive was built.

Solution/Action:

The six CPUs with removable drives and the supported images are listed in Table 1. To view the CPU Option, press **System, Show, System** on the analyzer. One of the first two options listed will be the CPU option (if the instrument software version is \geq A.02.06). Instrument software versions prior to A.02.00 only supported the N9010A-PC1 CPU, which did not have a removable drive.

Note that two of the CPUs support only the WinXP operating system, two others support either WinXP or Windows Embedded Standard 7 (“WES7”), and the most recent two CPUs support only WES7 or Win10

Note that disk drives are imaged with only one operating system, either WinXP, WES7 or Win10, not multiple operating systems.

Table 1. EXA CPUs with Removable Drives and Supported Images

CPU Option	CPU Type	Supported Images		
		WinXP	WES7	<u>Win10</u>
N9010A-PC2	Intel ® Core TM Duo T2500 @2.00 GHz	\geq A.02.06 to A.14.16	Not Supported	<u>Not Supported</u>
N9010A-PC3	Intel ® Celeron M 440 @1.86 GHz	\geq A.04.26 to A.14.16	Not Supported	<u>Not Supported</u>
N9010A-PC4	Intel ® i7 L 620 @2.00 GHz	\geq A.10.04 to A.14.16	\geq A.12.13	<u>Not Supported</u>
N9010A-PC5	Intel ® Celeron P4505 @1.87 GHz	\geq A.10.04 to A.14.16	\geq A.12.13	<u>Not Supported</u>
N9010A-PC6	Intel ® Core i7 – 3615QE CPU @ 2.30 GHz	Not Supported	\geq A.16.05	\geq <u>A.20.22</u>
N9010A-PC7	Intel ® Dual-Core Celeron ® CPU @ 2.2 GHz	Not Supported	\geq A.16.05	\geq <u>A.20.22</u>

To determine what the image version is for a particular disk drive, navigate to the C:\TEMP folder. In there you will find an Installer file, such as “XSA_Installer_A.12.13_x64.exe”. The “A.12.13.” portion of the filename indicates the image version.

The image version is the version of the XSA Signal Analyzer application. In the case of the example above, this would be A.12.13. According to Table 1, this version is compatible with N9010A-PC4 and N9010A-PC5, but not N9010A-PC6 or N9010A-PC7.

Even if a disk drive has a compatible image, this does not mean that the drive can be easily shared amongst several different types of CPUs. When a new disk drive is installed in an analyzer, the disk drive is “unsealed” at first power-up. As part of the unsealing process, the operating system discovers what CPU hardware is being used and configures itself properly for that CPU hardware.

Once the disk drive has been unsealed, it is configured only for the type of CPU in which it was originally installed.

If it is necessary to use the disk drive on a CPU of a different type (a “new CPU”), and the image on the disk drive can support the new CPU, it will be necessary to perform an Agilent Recovery or Instrument Image Recovery. This essentially returns the disk drive back to what it was like when it was new. Any updates to the XSA Signal Analyzer application and any user-installed software on the C: drive is overwritten by the recovery process. The recovery process, however, does not delete or overwrite any files on the D: (user documents) or E: (calibration data) drives.

As part of the recovery process, the operating system will again discover what CPU hardware is being used and configure itself appropriately.

CAUTION: Keysight does not support sharing a disk drive amongst several analyzers, even if the CPUs are all of the same type. The reason for this is that the disk drive also contains the calibration files which are unique for each analyzer. Sharing a drive amongst several analyzers requires separately backing up and archiving the calibration files for each analyzer and making sure these backups and archives are updated whenever the analyzer is calibrated or repaired. This requires great discipline.

An upgrade kit is available to upgrade to the N9010A-PC6 CPU. This kit, N9094AK-PC6, includes a new solid state drive (SSD) with the currently released instrument software installed. This kit is designed to upgrade the CPU on a properly-functioning analyzer and should not be used to replace a damaged CPU. The kit can also be used to upgrade the operating system from WinXP to WES7 or Win10 or from WES7 to Win10. Upgrades require a properly-functioning analyzer so that it will be possible to back-up the calibration files using the original CPU and then restore the calibration files once the new CPU is installed.

CPU replacement kits are also available to replace damaged CPUs. These kits do not include a new SSD since it should be possible to re-use the original SSD or HDD. These kits should not be used to upgrade from one CPU type to another. Refer also to service note N9010A-01C for CPU replacement kit part numbers.

Revision History:

Date	Service Note Revision	Author	Reason for Change
02 May 2012	03	BDT	As Published
21 Dec 2016	03A	BDT	Added reference to WES7 and PC6 and PC7
11 February 2019	03B	BDT	Added references to Win10 Operating System