

# INFORMATION ONLY

## N9030A-06

# S E R V I C E N O T E

Supersedes:  
NONE

## N9030A PXA Signal Analyzer

Serial Numbers: ALL

### A26 Wideband Digital IF Assembly Requires Instrument Software Upgrade

The A26 Wideband Digital IF Assembly, p/n N9020-60257, requires SW revision A.13.12 or later.

#### Parts Required:

P/N	Description	Qty.
NONE		

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:	
<b>INFORMATION ONLY</b>	
Calibration Required X Calibration NOT Required	PRODUCT LINE: 12 AUTHOR: BS
ADDITIONAL INFORMATION:	

© AGILENT TECHNOLOGIES, INC. 2014  
PRINTED IN U.S.A.

June 20, 2014

Rev. 22



**Situation:**

The A26 Wideband Digital IF Assembly, p/n N9020-60257 was introduced in June 2014, and because of some circuit changes, a minimum software revision is required when installing this assembly. This Wideband Digital IF Assembly is only present on instruments with option B1X or B85. Instrument specifications are unchanged.

**Solution/Action:**

The Instrument software must be revision A.13.12 or later for the Wideband Digital IF Assembly to function.

The latest N9030A PXA software can be down loaded for [www.agilent.com/find/xseries\\_software](http://www.agilent.com/find/xseries_software)

**Revision History:**

Revision Number	Date	Author	Reason For Change
1.0	6/20/2014	BS	As published