Modification Recommended Service Note

Supersedes: U1272A-07

U1272A Handheld Digital Multimeter Serial Numbers: MY00000000 / MY56999999

The Problem -Interference observed in the DC current measurement below 80MHz environment

Parts Required:

P/N Description Qty.

NONE

ADMINISTRATIVE INFORMATION

ACTION CATEGORY:	X ON SPECIFIED FAILURE [[]] AGREEABLE TIME	STANDARDS LABOR: 0.5 Hours	
LOCATION CATEGORY:	[[]] CUSTOMER INSTALLABLE [[]] ON-SITE (active On-site contract required) X SERVICE CENTER [[]] CHANNEL PARTNERS	INVENTORY: X SCRAP PARTS: [[]] RETURN []] SCRAP []] SEE TEXT
AVAILABILITY	: PRODUCT'S SUPPORT LIFE	NO CHARGE AVAILABLE UNTIL: 31-Jan-2018	3
	[[]] Calibration Required X Calibration NOT Required	PRODUCT LINE: GM AUTHOR: SW	

ADDITIONAL INFORMATION:



Situation:

The U1270 series handheld digital multimeter is susceptible to interference in DC current measurement below 80MHz environment. The symptom could be significantly improved by turning on the low pass filter function of the multimeter when performing DC current measurements. However, the filter design of the product did not sufficiently attenuate external noise/RF signal in the current measurement path. Customers may observe some random fluctuating readings in a noisy RF environment.

Solution/Action:

Keysight Technologies has carried out a product enhancement on the DC current measurement below 80MHz environment.

Customers who have concern on the interference observed in the DC current measurement below 80MHz environment, may contact Keysight Technologies' Customer Contact Center at www.keysight.com/find/contactus for repair services.

Revision History:

Date	Service Note Revision	Author	Reason for Change
6 Jan 2017	01	SW	As Published
13 Feb 2017	02	SW	Update Action Category to OSF