

U1620A-01

S E R V I C E N O T E

Supersedes:
NONE

U1620A 200MHz Handheld Digital Oscilloscope

Serial Numbers: MY52330000/ MY99999999

Power off sequence enhancement.

Parts Required:

P/N	Description	Qty.
NONE		

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:	
INFORMATION ONLY	
<input type="checkbox"/> Calibration Required	PRODUCT LINE: WC
<input checked="" type="checkbox"/> Calibration NOT Required	AUTHOR: SJ
ADDITIONAL INFORMATION:	

© AGILENT TECHNOLOGIES, INC. 2012
PRINTED IN U.S.A.

October 15, 2012




Rev. 20



Agilent Technologies

Situation:

There is a product enhancement for serial number starting with MY52330000, which allow better user experience during power off the unit.

Before Enhancement	After Enhancement
<p>To power off the unit, press and hold the  button for approximately 1 second.</p> <p>Unit will power off automatically after 1 second.</p>	<p>To power off the unit, press and hold the  button for approximately 1 second.</p> <p>A power off screen will pop up (as per image below) for user to select either “OK” by pressing F4 button to proceed for power off the unit, or “Cancel” by pressing F5 button to continue using the unit.</p> 

Solution/Action:

No action required.