

U1701B-01

# Information Only Service Note

Supersedes:  
NONE

## U1701B Handheld Capacitance Meter

Serial Numbers: ALL

**The Problem** – Accuracy specification for 1000pF is now typical.

### Parts Required:

P/N	Description	Qty.
-----	-------------	------

NONE

### ADMINISTRATIVE INFORMATION

---

Calibration Required  
 Calibration NOT Required

PRODUCT LINE: GM  
AUTHOR: WY

---

ADDITIONAL INFORMATION:

---

**Situation:**

It was recently discovered that the product is not able to establish calibration traceability with the recommended test methodology for 1000pF range.  
Hence, the accuracy specification for capacitance measurement at 1000pF range is now typical.  
No changes to the accuracy specification for the other ranges.

**Solution/Action:**

No action required.

**Revision History:**

Date	Service Note Revision	Author	Reason for Change
24 Jan 2022	01	WanYee	As Published

