

U1732C-01

Modification Recommended Service Note

Supersedes:
NONE

U1732C – Handheld LCR Meter

Serial Numbers: MY60299005 to MY63190070

The Problem – Units may show out-of-specification issues in specific measurements and parameters.

Parts Required:

P/N Description Qty.

NONE

ADMINISTRATIVE INFORMATION

ACTION	<input checked="" type="checkbox"/> ON SPECIFIED FAILURE	STANDARDS		
CATEGORY:	<input type="checkbox"/> AGREEABLE TIME	LABOR:	0.5 Hour	
LOCATION	<input checked="" type="checkbox"/> SERVICE CENTER	SERVICE:	<input type="checkbox"/> RETURN	USED <input type="checkbox"/> RETURN
CATEGORY:	<input type="checkbox"/> ON-SITE (active On-site contract required)	INVENTORY:	<input type="checkbox"/> SCRAP	PARTS: <input type="checkbox"/> SCRAP
	<input type="checkbox"/> CHANNEL PARTNERS		<input checked="" type="checkbox"/> SEE TEXT	<input checked="" type="checkbox"/> SEE TEXT
AVAILABILITY:	PRODUCT'S SUPPORT LIFE	NO CHARGE AVAILABLE UNTIL:	30 th September 2024	
	<input checked="" type="checkbox"/> Calibration Required	PRODUCT LINE:	GM	
	<input type="checkbox"/> Calibration NOT Required	AUTHOR:	ET	

ADDITIONAL INFORMATION:
No impact on service inventory

Situation

U1732C Handheld LCR meter may exhibit out-of-specification issues in specific measurements and parameters listed below,

Measurement	Range	Frequency
Capacitance	20 μ F	100 Hz
		120 Hz
Inductance	20 mH	1 kHz
	2000 mH	100 Hz
		120 Hz

Solution/Action:

U1732C Handheld LCR meter that exhibits out-of-specification issues as mentioned in this service note must be adjusted. Kindly send your unit to a Keysight Service Center for calibration adjustment.

Revision History:

Date	Service Note Revision	Author	Reason for Change
14 th Sep 2023	01	ET	As Published