

U8903A-09

S E R V I C E N O T E

Supersedes:
NONE

Agilent U8903A Audio Analyzer

Serial Numbers: MY52250005 to MY99999999

HP8903B Mode Disable at Default Setting

Parts Required:

P/N	Description	Qty.
NONE		

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:	
INFORMATION ONLY	
<input type="checkbox"/> Calibration Required	PRODUCT LINE: WC
<input checked="" type="checkbox"/> Calibration NOT Required	AUTHOR: CTK
ADDITIONAL INFORMATION:	

© AGILENT TECHNOLOGIES, INC. 2012
PRINTED IN U.S.A.

July 13, 2012

Rev. 20



Situation:

Starting from Serial Number above, U8903A will shipped with 'HP8903B mode' disabled as default setting. Most customers do not use the 'HP8903B mode', and would prefer to have it disabled.

Solution/Action:

'HP8903B mode' was enabled as default for firmware 2.10.3.0 and below. Customers can turn it off by using this sequence on the front panel : "System->HP8903B Config->Mode->Off".

U8903A firmware released after version 2.10.3.0, will have the 'HP8903B mode' disabled as default.